

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112119\
 Data File : VN059347.D
 Acq On : 21 Nov 2019 19:30
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Nov 22 07:16:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 21 08:44:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	369032	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	571255	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	478252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	177913	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	212514	48.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.00%	
35) Dibromofluoromethane	7.57	113	167937	49.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.00%	
50) Toluene-d8	10.08	98	682033	50.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.12%	
62) 4-Bromofluorobenzene	12.40	95	218096	45.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112119\
Data File : VN059347.D
Acq On : 21 Nov 2019 19:30
Operator : JC/SP
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Nov 22 07:16:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 21 08:44:45 2019
Response via : Initial Calibration

