

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112122\
 Data File : VN075345.D
 Acq On : 21 Nov 2022 15:15
 Operator : JC\MD
 Sample : VN1121MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1121MBL01

Quant Time: Nov 22 08:19:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 08:17:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	129635	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	221572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	226467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	119088	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	70146	100.804	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	201.600%#
35) Dibromofluoromethane	8.171	113	71919	82.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	165.260%#
50) Toluene-d8	10.565	98	237894	121.350	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	242.700%#
62) 4-Bromofluorobenzene	12.847	95	108872	75.240	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	150.480%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112122\
Data File : VN075345.D
Acq On : 21 Nov 2022 15:15
Operator : JC\MD
Sample : VN1121MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1121MBL01

Quant Time: Nov 22 08:19:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 08:17:11 2022
Response via : Initial Calibration

