

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112122\
 Data File : VN075349.D
 Acq On : 21 Nov 2022 18:18
 Operator : JC\MD
 Sample : N5699-01ME
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AREA-COMP-EXCAV-1ST-AVE

Quant Time: Nov 22 08:19:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 08:17:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	295546	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	504148	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	425656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	181109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	57166	36.034	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	72.060%#
35) Dibromofluoromethane	8.165	113	68931	34.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	69.620%#
50) Toluene-d8	10.565	98	156320	35.045	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	70.100%#
62) 4-Bromofluorobenzene	12.847	95	100738	30.597	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	61.200%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112122\
Data File : VN075349.D
Acq On : 21 Nov 2022 18:18
Operator : JC\MD
Sample : N5699-01ME
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AREA-COMP-EXCAV-1ST-AVE

Quant Time: Nov 22 08:19:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 08:17:11 2022
Response via : Initial Calibration

