

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112122\
 Data File : VN075344.D
 Acq On : 21 Nov 2022 14:35
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN112122

Quant Time: Nov 23 00:20:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.382	0.370	3.1	100	0.00
3 P	Chloromethane	0.476	0.447	6.1	104	0.00
4 C	Vinyl Chloride	0.600	0.585	2.5#	103	0.00
5 T	Bromomethane	0.527	0.506	4.0	94	0.01
6 T	Chloroethane	0.454	0.440	3.1	106	0.00
7 T	Trichlorofluoromethane	0.857	0.828	3.4	102	0.00
8 T	Diethyl Ether	0.300	0.307	-2.3	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.484	-2.3	106	0.00
10 T	Methyl Iodide	0.766	0.758	1.0	102	0.00
11 T	Tert butyl alcohol	0.113	0.106	6.2	105	0.00
12 CM	1,1-Dichloroethene	0.463	0.468	-1.1#	105	0.01
13 T	Acrolein	0.098	0.095	3.1	104	0.00
14 T	Allyl chloride	0.599	0.560	6.5	104	0.00
15 T	Acrylonitrile	0.258	0.252	2.3	102	0.00
16 T	Acetone	0.250	0.248	0.8	109	0.00
17 T	Carbon Disulfide	1.203	1.178	2.1	102	0.00
18 T	Methyl Acetate	0.741	0.642	13.4	103	0.00
19 T	Methyl tert-butyl Ether	1.640	1.624	1.0	104	0.00
20 T	Methylene Chloride	0.587	0.528	10.1	104	0.00
21 T	trans-1,2-Dichloroethene	0.528	0.516	2.3	104	0.00
22 T	Diisopropyl ether	1.367	1.348	1.4	101	0.00
23 T	Vinyl Acetate	1.087	1.108	-1.9	101	0.00
24 P	1,1-Dichloroethane	0.877	0.882	-0.6	103	0.00
25 T	2-Butanone	0.350	0.335	4.3	103	0.00
26 T	2,2-Dichloropropane	0.815	0.791	2.9	100	0.00
27 T	cis-1,2-Dichloroethene	0.597	0.605	-1.3	101	0.00
28 T	Bromochloromethane	0.356	0.359	-0.8	106	0.00
29 T	Tetrahydrofuran	0.213	0.212	0.5	103	0.00
30 C	Chloroform	1.007	0.962	4.5#	102	0.00
31 T	Cyclohexane	0.853	0.786	7.9	101	0.00
32 T	1,1,1-Trichloroethane	0.915	0.898	1.9	101	0.00
33 S	1,2-Dichloroethane-d4	0.268	0.263	1.9	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.196	0.201	-2.6	99	0.00
36 T	1,1-Dichloropropene	0.428	0.437	-2.1	102	0.00
37 T	Ethyl Acetate	0.386	0.396	-2.6	99	0.00
38 T	Carbon Tetrachloride	0.471	0.495	-5.1	102	0.00
39 T	Methylcyclohexane	0.536	0.578	-7.8	106	0.00
40 TM	Benzene	1.311	1.317	-0.5	101	0.00
41 T	Methacrylonitrile	0.197	0.213	-8.1	111	0.00
42 TM	1,2-Dichloroethane	0.439	0.439	0.0	102	0.00
43 T	Isopropyl Acetate	0.619	0.653	-5.5	104	0.00
44 TM	Trichloroethene	0.359	0.365	-1.7	104	0.00
45 C	1,2-Dichloropropane	0.313	0.320	-2.2#	103	0.00
46 T	Dibromomethane	0.236	0.242	-2.5	103	0.00
47 T	Bromodichloromethane	0.458	0.475	-3.7	103	0.00
48 T	Methyl methacrylate	0.286	0.296	-3.5	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112122\
 Data File : VN075344.D
 Acq On : 21 Nov 2022 14:35
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN112122

Quant Time: Nov 23 00:20:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	101	0.00
50 S	Toluene-d8	0.442	0.461	-4.3	100	0.00
51 T	4-Methyl-2-Pentanone	0.394	0.407	-3.3	103	0.00
52 CM	Toluene	0.845	0.899	-6.4#	104	0.00
53 T	t-1,3-Dichloropropene	0.460	0.508	-10.4	102	0.00
54 T	cis-1,3-Dichloropropene	0.499	0.550	-10.2	105	0.00
55 T	1,1,2-Trichloroethane	0.344	0.348	-1.2	104	0.00
56 T	Ethyl methacrylate	0.465	0.519	-11.6	103	0.00
57 T	1,3-Dichloropropane	0.561	0.566	-0.9	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.168	0.193	-14.9	105	0.00
59 T	2-Hexanone	0.296	0.314	-6.1	103	0.00
60 T	Dibromochloromethane	0.353	0.396	-12.2	101	0.00
61 T	1,2-Dibromoethane	0.357	0.369	-3.4	102	0.00
62 S	4-Bromofluorobenzene	0.327	0.349	-6.7	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.382	0.364	4.7	103	0.00
65 PM	Chlorobenzene	1.033	1.026	0.7	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.390	-1.3	101	0.00
67 C	Ethyl Benzene	1.762	1.849	-4.9#	103	0.00
68 T	m/p-Xylenes	0.715	0.741	-3.6	103	0.00
69 T	o-Xylene	0.711	0.731	-2.8	104	0.00
70 T	Styrene	1.121	1.206	-7.6	103	0.00
71 P	Bromoform	0.293	0.329	-12.3	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.653	3.709	-1.5	104	0.00
74 T	N-amyl acetate	1.123	1.194	-6.3	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.114	1.078	3.2	103	0.00
76 T	1,2,3-Trichloropropane	0.945	0.776	17.9	87	0.00
77 T	Bromobenzene	0.902	0.898	0.4	103	0.00
78 T	n-propylbenzene	4.077	4.235	-3.9	103	0.00
79 T	2-Chlorotoluene	2.456	2.502	-1.9	104	0.00
80 T	1,3,5-Trimethylbenzene	3.035	3.180	-4.8	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.320	0.342	-6.9	102	0.00
82 T	4-Chlorotoluene	2.456	2.502	-1.9	104	0.00
83 T	tert-Butylbenzene	2.767	2.795	-1.0	102	0.00
84 T	1,2,4-Trimethylbenzene	3.044	3.241	-6.5	105	0.00
85 T	sec-Butylbenzene	3.755	4.012	-6.8	103	0.00
86 T	p-Isopropyltoluene	3.159	3.433	-8.7	103	0.00
87 T	1,3-Dichlorobenzene	1.709	1.741	-1.9	103	0.00
88 T	1,4-Dichlorobenzene	1.736	1.738	-0.1	104	0.00
89 T	n-Butylbenzene	2.668	2.931	-9.9	105	0.00
90 T	Hexachloroethane	0.545	0.576	-5.7	100	0.00
91 T	1,2-Dichlorobenzene	1.666	1.687	-1.3	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.211	0.224	-6.2	102	0.00
93 T	1,2,4-Trichlorobenzene	0.854	0.959	-12.3	102	0.00
94 T	Hexachlorobutadiene	0.445	0.494	-11.0	105	0.00
95 T	Naphthalene	2.685	2.912	-8.5	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112122\
Data File : VN075344.D
Acq On : 21 Nov 2022 14:35
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN112122

Quant Time: Nov 23 00:20:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.839	0.913	-8.8	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6