

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112123\
 Data File : VN080234.D
 Acq On : 21 Nov 2023 16:40
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 22 00:08:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.745	0.745	0.0	84	0.00
3 P	Chloromethane	0.560	0.452	19.3	70	0.00
4 C	Vinyl Chloride	0.593	0.536	9.6#	77	0.00
5 T	Bromomethane	0.402	0.352	12.4	78	0.00
6 T	Chloroethane	0.447	0.330	26.2#	75	0.00
7 T	Trichlorofluoromethane	1.236	1.237	-0.1	84	0.00
8 T	Diethyl Ether	0.360	0.324	10.0	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.565	0.554	1.9	84	0.00
10 T	Methyl Iodide	0.611	0.633	-3.6	82	0.00
11 T	Tert butyl alcohol	0.091	0.079	13.2	73	0.00
12 CM	1,1-Dichloroethene	0.528	0.503	4.7#	81	0.00
13 T	Acrolein	0.080	0.072	10.0	82	0.00
14 T	Allyl chloride	0.665	0.604	9.2	76	0.00
15 T	Acrylonitrile	0.244	0.208	14.8	73	0.00
16 T	Acetone	0.277	0.186	32.9#	65	0.00
17 T	Carbon Disulfide	1.578	1.316	16.6	74	0.00
18 T	Methyl Acetate	0.891	0.719	19.3	74	0.00
19 T	Methyl tert-butyl Ether	1.866	1.868	-0.1	83	0.00
20 T	Methylene Chloride	0.581	0.569	2.1	82	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.550	5.3	81	0.00
22 T	Diisopropyl ether	1.471	1.382	6.1	78	0.00
23 T	Vinyl Acetate	0.884	0.787	11.0	72	0.00
24 P	1,1-Dichloroethane	1.071	0.994	7.2	82	0.00
25 T	2-Butanone	0.329	0.241	26.7#	67	0.00
26 T	2,2-Dichloropropane	1.063	1.092	-2.7	86	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.631	4.1	82	0.00
28 T	Bromochloromethane	0.421	0.386	8.3	83	0.00
29 T	Tetrahydrofuran	0.177	0.146	17.5	68	0.00
30 C	Chloroform	1.198	1.182	1.3#	85	0.00
31 T	Cyclohexane	0.898	0.731	18.6	71	0.00
32 T	1,1,1-Trichloroethane	1.168	1.205	-3.2	87	0.00
33 S	1,2-Dichloroethane-d4	0.826	0.823	0.4	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.369	0.391	-6.0	91	0.00
36 T	1,1-Dichloropropene	0.534	0.558	-4.5	82	0.00
37 T	Ethyl Acetate	0.371	0.328	11.6	72	0.00
38 T	Carbon Tetrachloride	0.678	0.756	-11.5	88	0.00
39 T	Methylcyclohexane	0.515	0.533	-3.5	77	0.00
40 TM	Benzene	1.518	1.512	0.4	81	0.00
41 T	Methacrylonitrile	0.203	0.196	3.4	76	0.00
42 TM	1,2-Dichloroethane	0.621	0.682	-9.8	88	0.00
43 T	Isopropyl Acetate	0.721	0.605	16.1	72	0.00
44 TM	Trichloroethene	0.398	0.406	-2.0	84	0.00
45 C	1,2-Dichloropropane	0.353	0.347	1.7#	79	0.00
46 T	Dibromomethane	0.267	0.278	-4.1	84	0.00
47 T	Bromodichloromethane	0.587	0.646	-10.1	87	0.00
48 T	Methyl methacrylate	0.372	0.362	2.7	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112123\
 Data File : VN080234.D
 Acq On : 21 Nov 2023 16:40
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 22 00:08:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.005	16.7	71	0.00
50 S	Toluene-d8	1.330	1.380	-3.8	85	0.00
51 T	4-Methyl-2-Pentanone	0.373	0.342	8.3	72	0.00
52 CM	Toluene	0.950	0.988	-4.0#	82	0.00
53 T	t-1,3-Dichloropropene	0.604	0.657	-8.8	84	0.00
54 T	cis-1,3-Dichloropropene	0.617	0.660	-7.0	82	0.00
55 T	1,1,2-Trichloroethane	0.363	0.365	-0.6	83	0.00
56 T	Ethyl methacrylate	0.370	0.397	-7.3	77	0.00
57 T	1,3-Dichloropropane	0.633	0.644	-1.7	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.209	0.237	-13.4	94	0.00
59 T	2-Hexanone	0.300	0.263	12.3	71	0.00
60 T	Dibromochloromethane	0.434	0.479	-10.4	87	0.00
61 T	1,2-Dibromoethane	0.370	0.379	-2.4	83	0.00
62 S	4-Bromofluorobenzene	0.502	0.532	-6.0	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.382	0.412	-7.9	91	0.00
65 PM	Chlorobenzene	1.133	1.106	2.4	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.440	0.479	-8.9	87	0.00
67 C	Ethyl Benzene	2.016	2.149	-6.6#	85	0.00
68 T	m/p-Xylenes	0.735	0.785	-6.8	83	0.00
69 T	o-Xylene	0.693	0.750	-8.2	84	0.00
70 T	Styrene	1.031	1.118	-8.4	83	0.00
71 P	Bromoform	0.326	0.348	-6.7	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.885	4.078	-5.0	84	0.00
74 T	N-amyl acetate	1.212	1.099	9.3	73	0.00
75 P	1,1,2,2-Tetrachloroethane	1.163	0.994	14.5	76	0.00
76 T	1,2,3-Trichloropropane	1.225	1.107	9.6	77	0.00
77 T	Bromobenzene	0.980	0.989	-0.9	85	0.00
78 T	n-propylbenzene	4.345	4.559	-4.9	83	0.00
79 T	2-Chlorotoluene	2.822	2.872	-1.8	84	0.00
80 T	1,3,5-Trimethylbenzene	3.382	3.609	-6.7	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.363	0.321	11.6	72	0.00
82 T	4-Chlorotoluene	2.891	2.973	-2.8	85	0.00
83 T	tert-Butylbenzene	2.712	2.889	-6.5	84	0.00
84 T	1,2,4-Trimethylbenzene	3.330	3.566	-7.1	85	0.00
85 T	sec-Butylbenzene	3.431	3.694	-7.7	84	0.00
86 T	p-Isopropyltoluene	3.065	3.418	-11.5	87	0.00
87 T	1,3-Dichlorobenzene	1.765	1.726	2.2	85	0.00
88 T	1,4-Dichlorobenzene	1.774	1.727	2.6	86	0.00
89 T	n-Butylbenzene	2.565	2.771	-8.0	83	0.00
90 T	Hexachloroethane	0.598	0.583	2.5	82	0.00
91 T	1,2-Dichlorobenzene	1.689	1.662	1.6	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.236	12.6	78	0.00
93 T	1,2,4-Trichlorobenzene	0.982	0.967	1.5	86	0.00
94 T	Hexachlorobutadiene	0.565	0.553	2.1	88	0.00
95 T	Naphthalene	2.848	2.471	13.2	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112123\
Data File : VN080234.D
Acq On : 21 Nov 2023 16:40
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 22 00:08:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.921	0.885	3.9	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6