

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112124\
 Data File : VN084990.D
 Acq On : 21 Nov 2024 11:43
 Operator : JC\MD
 Sample : VN1121WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1121WBL01

Quant Time: Nov 21 14:43:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 01 02:38:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	26986	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	144485	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	123949	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	67156	30.954	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.167%
60) 4-Bromofluorobenzene	12.847	95	52082	24.404	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.333%
63) Toluene-d8	10.565	98	164490	28.033	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	93.433%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112124\
Data File : VN084990.D
Acq On : 21 Nov 2024 11:43
Operator : JC\MD
Sample : VN1121WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1121WBL01

Quant Time: Nov 21 14:43:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 01 02:38:04 2024
Response via : Initial Calibration

