

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112124\
 Data File : VN084996.D
 Acq On : 21 Nov 2024 15:57
 Operator : JC\MD
 Sample : VN1121MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1121MBL01

Quant Time: Nov 22 04:34:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	156478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	281110	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241543	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118972	52.641	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	105.280%		
35) Dibromofluoromethane	8.165	113	95027	49.941	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.880%		
50) Toluene-d8	10.565	98	321033	45.802	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.600%		
62) 4-Bromofluorobenzene	12.847	95	116569	44.499	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	89.000%		
Target Compounds						
58) 2-Chloroethyl Vinyl ether	10.159	63	1617	1.280	ug/l #	86
82) 4-Chlorotoluene	13.223	91	6918	1.210	ug/l	96
87) 1,3-Dichlorobenzene	13.729	146	5417	1.455	ug/l	96
88) 1,4-Dichlorobenzene	13.812	146	6995	1.205	ug/l	86
89) n-Butylbenzene	14.059	91	6706	1.271	ug/l	96
91) 1,2-Dichlorobenzene	14.106	146	4267	1.192	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.717	75	487	1.015	ug/l #	54
93) 1,2,4-Trichlorobenzene	15.394	180	4626	2.362	ug/l	99
94) Hexachlorobutadiene	15.494	225	1588	1.846	ug/l	94
95) Naphthalene	15.647	128	12828	2.188	ug/l #	93
96) 1,2,3-Trichlorobenzene	15.841	180	4578	2.408	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112124\
Data File : VN084996.D
Acq On : 21 Nov 2024 15:57
Operator : JC\MD
Sample : VN1121MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1121MBL01

Quant Time: Nov 22 04:34:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

