

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112124\
 Data File : VN085000.D
 Acq On : 21 Nov 2024 17:44
 Operator : JC\MD
 Sample : P4929-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ARS520

Quant Time: Nov 22 04:37:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	283234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	106885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	115812	51.558	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.120%			
35) Dibromofluoromethane	8.165	113	91628	47.794	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.580%			
50) Toluene-d8	10.565	98	328925	46.576	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 93.160%			
62) 4-Bromofluorobenzene	12.847	95	120165	45.528	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 91.060%			
Target Compounds						
16) Acetone	4.430	43	76044	115.663	ug/l	94
20) Methylene Chloride	5.253	84	2002	1.013	ug/l #	62
43) Isopropyl Acetate	8.688	43	95775	20.929	ug/l #	75
93) 1,2,4-Trichlorobenzene	15.394	180	2325	1.125	ug/l	96
95) Naphthalene	15.641	128	8455	1.367	ug/l #	94
96) 1,2,3-Trichlorobenzene	15.835	180	2280	1.136	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112124\
Data File : VN085000.D
Acq On : 21 Nov 2024 17:44
Operator : JC\MD
Sample : P4929-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ARS520

Quant Time: Nov 22 04:37:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

