

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112124\
 Data File : VN085001.D
 Acq On : 21 Nov 2024 18:08
 Operator : JC\MD
 Sample : P4892-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WB-310-BOT

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/22/2024
 Supervised By : Mahesh Dadoda 11/22/2024

Quant Time: Nov 22 04:38:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	163820	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	297677	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111856	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121255	51.246	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.500%			
35) Dibromofluoromethane	8.165	113	95988	47.639	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.280%			
50) Toluene-d8	10.565	98	346670	46.707	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 93.420%			
62) 4-Bromofluorobenzene	12.847	95	128097	46.178	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 92.360%			
Target Compounds						
16) Acetone	4.436	43	21682	29.204	ug/l	Qvalue 100
43) Isopropyl Acetate	8.688	43	143683m	30.295	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112124\
Data File : VN085001.D
Acq On : 21 Nov 2024 18:08
Operator : JC\MD
Sample : P4892-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WB-310-BOT

Quant Time: Nov 22 04:38:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/22/2024
Supervised By :Mahesh Dadoda 11/22/2024

