

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112219\
 Data File : VN059357.D
 Acq On : 22 Nov 2019 12:29
 Operator : JC/SP
 Sample : K5934-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Nov 22 13:44:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 21 08:44:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	365523	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	560156	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	485752	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	180416	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	206689	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
35) Dibromofluoromethane	7.57	113	166089	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
50) Toluene-d8	10.08	98	678963	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	12.40	95	216297	46.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.60%	
Target Compounds						
16) Acetone	3.79	43	19349	9.840	ug/l	97
20) Methylene Chloride	4.53	84	501285	113.028	ug/l	100
43) Isopropyl Acetate	8.15	43	94099	9.907	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112219\
Data File : VN059357.D
Acq On : 22 Nov 2019 12:29
Operator : JC/SP
Sample : K5934-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Nov 22 13:44:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 21 08:44:45 2019
Response via : Initial Calibration

