

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112219\
 Data File : VN059375.D
 Acq On : 22 Nov 2019 19:04
 Operator : JC/SP
 Sample : K5676-25
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-112019

Quant Time: Nov 22 23:02:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 21 08:44:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	362826	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	557258	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	485705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	191882	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	208401	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
35) Dibromofluoromethane	7.57	113	166389	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
50) Toluene-d8	10.08	98	683894	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	12.40	95	222073	47.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.58%	
Target Compounds						
16) Acetone	3.80	43	10453	5.355	ug/l	99
20) Methylene Chloride	4.53	84	44782	10.172	ug/l	96
43) Isopropyl Acetate	8.15	43	76751	8.123	ug/l #	92
84) 1,2,4-Trimethylbenzene	13.04	105	18137	1.401	ug/l	97
95) Naphthalene	15.13	128	136551	14.474	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112219\
Data File : VN059375.D
Acq On : 22 Nov 2019 19:04
Operator : JC/SP
Sample : K5676-25
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-112019

Quant Time: Nov 22 23:02:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 21 08:44:45 2019
Response via : Initial Calibration

