

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112219\
 Data File : VN059377.D
 Acq On : 22 Nov 2019 19:48
 Operator : JC/SP
 Sample : PB125021TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB125021TB

Quant Time: Nov 22 23:08:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 21 08:44:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	354237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	540355	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	472446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	182531	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	202168	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
35) Dibromofluoromethane	7.57	113	162258	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
50) Toluene-d8	10.08	98	655356	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	12.40	95	213628	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
Target Compounds						
95) Naphthalene	15.13	128	8262	3.887	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112219\
Data File : VN059377.D
Acq On : 22 Nov 2019 19:48
Operator : JC/SP
Sample : PB125021TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB125021TB

Quant Time: Nov 22 23:08:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 21 08:44:45 2019
Response via : Initial Calibration

