

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112219\
 Data File : VN059361.D
 Acq On : 22 Nov 2019 13:57
 Operator : JC/SP
 Sample : K5957-11
 Misc : 2.53g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5-(12-14)

Quant Time: Nov 27 04:22:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 21 08:44:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	342928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	528077	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	455643	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	167596	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	197728	48.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.12%	
35) Dibromofluoromethane	7.57	113	154883	48.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.78%	
50) Toluene-d8	10.08	98	637947	50.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.30%	
62) 4-Bromofluorobenzene	12.40	95	203809	46.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.56%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112219\
Data File : VN059361.D
Acq On : 22 Nov 2019 13:57
Operator : JC/SP
Sample : K5957-11
Misc : 2.53g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5-(12-14)

Quant Time: Nov 27 04:22:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 21 08:44:45 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1826

Delta R.T. -0.00 min

Lab File: VN059361.D

Acq: 22 Nov 2019 13:57

Instrument :

MSVOA_N

ClientSampleId :


MW-5-(12-14)

Tgt Ion:168 Resp: 342928

Ion Ratio Lower Upper

168 100

99 54.6 44.6 66.8



Abundance Ion 167.90 (167.60 to 168.60): VN059361.D (-1734) (-)

Ion 98.90 (98.60 to 99.60): VN059361.D (-1734) (-)

7.64

150000



100000

50000

0

7.50 7.60 7.70 7.80

Time-->



#33

1,2-Dichloroethane-d4

Concen: 48.055 ug/l

RT: 8.01 min Scan# 1939

Delta R.T. 0.00 min

Lab File: VN059361.D


Acq: 22 Nov 2019 13:57

Tgt Ion: 65 Resp: 197728

Ion Ratio Lower Upper

65 100

67 54.2 0.0 109.4



Abundance Ion 64.90 (64.60 to 65.60): VN059361.D (-1846) (-)

Ion 66.90 (66.60 to 67.60): VN059361.D (-1846) (-)

8.01

100000

80000

60000


40000


20000

0

7.90 8.00 8.10

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2114

Delta R.T. 0.00 min

Lab File: VN059361.D

Acq: 22 Nov 2019 13:57

Instrument :

MSVOA_N

ClientSampleId :

MW-5-(12-14)


Tgt Ion:114 Resp: 528077

Ion Ratio Lower Upper

114 100

63 21.3 0.0 42.2




88 15.9 0.0 32.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 48.887 ug/l

RT: 7.57 min Scan# 1803

Delta R.T. 0.00 min

Lab File: VN059361.D

Acq: 22 Nov 2019 13:57


Tgt Ion:113 Resp: 154883

Ion Ratio Lower Upper

113 100

111 103.3 82.7 124.1



192 20.7 16.3 24.5




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.651 ug/l

RT: 10.08 min Scan# 2584

Delta R.T. 0.00 min

Lab File: VN059361.D

Acq: 22 Nov 2019 13:57

Instrument :

MSVOA_N

ClientSampleId :





MW-5-(12-14)

Tgt Ion: 98 Resp: 637947

Ion Ratio Lower Upper

98 100

100 65.0 51.7 77.5



#62

4-Bromofluorobenzene

Concen: 46.280 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN059361.D

Acq: 22 Nov 2019 13:57




Tgt Ion: 95 Resp: 203809


Ion Ratio Lower Upper

95 100

174 78.8 0.0 157.4

176 76.8 0.0 151.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN059361.D

Acq: 22 Nov 2019 13:57

Instrument :

MSVOA_N

ClientSampleId :

MW-5-(12-14)


Tgt Ion:117 Resp: 455643

Ion Ratio Lower Upper

117 100

82 56.7 45.7 68.5




119 32.0 25.8 38.6



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN059361.D

Acq: 22 Nov 2019 13:57


Tgt Ion:152 Resp: 167596

Ion Ratio Lower Upper

152 100

115 58.9 29.3 88.0

150 156.5 0.0 346.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

