

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112221\
 Data File : VN069617.D
 Acq On : 22 Nov 2021 18:19
 Operator : JC/MD
 Sample : M4806-03 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-3

Quant Time: Nov 23 03:54:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111021W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Nov 10 14:36:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	159735	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	858398	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.746	117	738390	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	434963	31.397	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.667%
60) 4-Bromofluorobenzene	12.733	95	277597	23.493	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	78.300%
63) Toluene-d8	10.443	98	1056103	30.977	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.267%
Target Compounds						
						Qvalue
25) Chloroform	7.820	83	982677	51.531	ug/l	99
41) Bromodichloromethane	9.764	83	59950	4.519	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112221\
 Data File : VN069617.D
 Acq On : 22 Nov 2021 18:19
 Operator : JC/MD
 Sample : M4806-03 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-3

Quant Time: Nov 23 03:54:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111021W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Nov 10 14:36:17 2021
 Response via : Initial Calibration

