

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075361.D
 Acq On : 22 Nov 2022 12:33
 Operator : JC\MD
 Sample : N5686-01ME
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-2-111722ME

Quant Time: Nov 23 07:06:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	334279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	576928	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	504027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	232693	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	93546	52.133	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.260%
35) Dibromofluoromethane	8.171	113	111519	49.206	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.420%
50) Toluene-d8	10.565	98	253675	49.697	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.847	95	186281	49.442	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075361.D
Acq On : 22 Nov 2022 12:33
Operator : JC\MD
Sample : N5686-01ME
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-2-111722ME

Quant Time: Nov 23 07:06:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

