

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075364.D
 Acq On : 22 Nov 2022 13:44
 Operator : JC\MD
 Sample : N5686-01
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-2-111722

Quant Time: Nov 22 14:05:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 08:17:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	287245	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101968	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	43243	46.398	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.800%
35) Dibromofluoromethane	8.171	113	51687	45.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.620%
50) Toluene-d8	10.565	98	115352	45.388	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.780%
62) 4-Bromofluorobenzene	12.853	95	77336	41.227	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.460%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075364.D
Acq On : 22 Nov 2022 13:44
Operator : JC\MD
Sample : N5686-01
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-2-111722

Quant Time: Nov 22 14:05:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 08:17:11 2022
Response via : Initial Calibration

