

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075369.D
 Acq On : 22 Nov 2022 15:44
 Operator : JC\MD
 Sample : PB149226ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149226ZHE#01

Quant Time: Nov 23 07:10:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	151511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	264100	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	243592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	102823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	42476	52.227	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.460%	
35) Dibromofluoromethane	8.171	113	51147	49.300	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.600%	
50) Toluene-d8	10.571	98	119136	50.985	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.980%	
62) 4-Bromofluorobenzene	12.853	95	84368	48.917	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.840%	
Target Compounds						
						Qvalue
8) Diethyl Ether	3.971	74	3226	3.546	ug/l	86
16) Acetone	4.442	43	65338	86.130	ug/l	97
20) Methylene Chloride	5.289	84	17415	10.306	ug/l #	86
37) Ethyl Acetate	7.565	43	22679	11.132	ug/l	98
43) Isopropyl Acetate	8.694	43	113407	34.681	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075369.D
Acq On : 22 Nov 2022 15:44
Operator : JC\MD
Sample : PB149226ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149226ZHE#01

Quant Time: Nov 23 07:10:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

