

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075370.D
 Acq On : 22 Nov 2022 16:08
 Operator : JC\MD
 Sample : PB149226ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149226ZHE#02

Quant Time: Nov 23 07:11:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	142434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	249802	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	232357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	41692	54.530	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.060%
35) Dibromofluoromethane	8.171	113	49463	50.405	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.820%
50) Toluene-d8	10.571	98	112149	50.742	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.480%
62) 4-Bromofluorobenzene	12.853	95	80360	49.260	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.520%
Target Compounds						
						Qvalue
8) Diethyl Ether	3.977	74	3122	3.651	ug/l	100
16) Acetone	4.441	43	65273	91.527	ug/l	100
20) Methylene Chloride	5.283	84	16710	10.546	ug/l	95
37) Ethyl Acetate	7.571	43	21329	11.069	ug/l	99
43) Isopropyl Acetate	8.694	43	125625	40.617	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075370.D
Acq On : 22 Nov 2022 16:08
Operator : JC\MD
Sample : PB149226ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149226ZHE#02

Quant Time: Nov 23 07:11:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

