

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075372.D
 Acq On : 22 Nov 2022 16:55
 Operator : JC\MD
 Sample : PB149226ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149226ZHE#04

Quant Time: Nov 23 07:11:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	140603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	254687	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	233392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	100408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	42402	56.181	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.360%
35) Dibromofluoromethane	8.171	113	50448	50.423	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.840%
50) Toluene-d8	10.571	98	115372	51.199	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.400%
62) 4-Bromofluorobenzene	12.853	95	81142	48.785	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.580%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.977	74	3010	3.566	ug/l	91
16) Acetone	4.442	43	64521	91.651	ug/l	89
20) Methylene Chloride	5.283	84	16751	10.730	ug/l	89
43) Isopropyl Acetate	8.694	43	156374	49.589	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075372.D
Acq On : 22 Nov 2022 16:55
Operator : JC\MD
Sample : PB149226ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149226ZHE#04

Quant Time: Nov 23 07:11:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

