

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075373.D
 Acq On : 22 Nov 2022 17:19
 Operator : JC\MD
 Sample : PB149226ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149226ZHE#05

Quant Time: Nov 23 07:12:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	140411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	250129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	226089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	95007	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	39791	52.794	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.580%
35) Dibromofluoromethane	8.171	113	48083	48.935	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	10.565	98	110808	50.070	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.140%
62) 4-Bromofluorobenzene	12.847	95	76756	46.989	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.980%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.977	74	2955	3.505	ug/l	91
16) Acetone	4.447	43	63409	90.195	ug/l	95
20) Methylene Chloride	5.289	84	16713	10.719	ug/l	90
37) Ethyl Acetate	7.571	43	21235	11.006	ug/l	97
43) Isopropyl Acetate	8.694	43	153512	49.568	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075373.D
Acq On : 22 Nov 2022 17:19
Operator : JC\MD
Sample : PB149226ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149226ZHE#05

Quant Time: Nov 23 07:12:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

