

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075379.D
 Acq On : 22 Nov 2022 19:43
 Operator : JC\MD
 Sample : N5749-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-82

Quant Time: Nov 23 07:15:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	160574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	278103	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	42761	49.610	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.220%
35) Dibromofluoromethane	8.171	113	52566	48.116	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.240%
50) Toluene-d8	10.571	98	117917	47.923	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.840%
62) 4-Bromofluorobenzene	12.853	95	86517	47.637	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.280%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	16796	20.891	ug/l	89
20) Methylene Chloride	5.289	84	6869	3.011	ug/l	92
37) Ethyl Acetate	7.571	43	11874	5.535	ug/l	96
43) Isopropyl Acetate	8.694	43	72554	21.071	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075379.D
Acq On : 22 Nov 2022 19:43
Operator : JC\MD
Sample : N5749-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-82

Quant Time: Nov 23 07:15:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

