

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075380.D
 Acq On : 22 Nov 2022 20:06
 Operator : JC\MD
 Sample : N5749-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-83

Quant Time: Nov 23 07:15:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	156736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	269272	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	236383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	43279	51.441	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.880%	
35) Dibromofluoromethane	8.177	113	52836	49.949	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.900%	
50) Toluene-d8	10.565	98	117817	49.452	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.900%	
62) 4-Bromofluorobenzene	12.847	95	84089	47.819	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.640%	
Target Compounds						
						Qvalue
16) Acetone	4.453	43	34691	44.206	ug/l	99
20) Methylene Chloride	5.289	84	7199	3.330	ug/l	83
43) Isopropyl Acetate	8.700	43	52692	15.804	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075380.D
Acq On : 22 Nov 2022 20:06
Operator : JC\MD
Sample : N5749-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-83

Quant Time: Nov 23 07:15:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

