

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
 Data File : VN075381.D
 Acq On : 22 Nov 2022 20:30
 Operator : JC\MD
 Sample : N5749-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-83.5

Quant Time: Nov 23 07:16:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	150289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	265525	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	235486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	102132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	43409	53.809	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.620%	
35) Dibromofluoromethane	8.177	113	53105	50.912	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.820%	
50) Toluene-d8	10.571	98	119863	51.021	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.040%	
62) 4-Bromofluorobenzene	12.847	95	83612	48.218	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.440%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	34861	46.328	ug/l	99
20) Methylene Chloride	5.294	84	7138	3.488	ug/l #	79
43) Isopropyl Acetate	8.694	43	47282	14.382	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112222\
Data File : VN075381.D
Acq On : 22 Nov 2022 20:30
Operator : JC\MD
Sample : N5749-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-83.5

Quant Time: Nov 23 07:16:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

