

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
 Data File : VN080237.D
 Acq On : 22 Nov 2023 10:45
 Operator : JC\MD
 Sample : VN1122MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1122MBL01

Quant Time: Nov 23 00:15:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	144057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	262905	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	276002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	113363	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	135880	57.126	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.260%
35) Dibromofluoromethane	8.171	113	104806	54.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.060%
50) Toluene-d8	10.571	98	345434	49.398	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.800%
62) 4-Bromofluorobenzene	12.853	95	129713	49.109	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.220%
Target Compounds						
79) 2-Chlorotoluene	13.223	91	1536	0.240	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
Data File : VN080237.D
Acq On : 22 Nov 2023 10:45
Operator : JC\MD
Sample : VN1122MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1122MBL01

Quant Time: Nov 23 00:15:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

