

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
 Data File : VN080238.D
 Acq On : 22 Nov 2023 11:09
 Operator : JC\MD
 Sample : VN1122WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1122WBL01

Quant Time: Nov 23 00:15:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	153572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	281247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	297202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127802	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	145249	57.281	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.560%
35) Dibromofluoromethane	8.171	113	112054	53.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.000%
50) Toluene-d8	10.571	98	373655	49.949	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.900%
62) 4-Bromofluorobenzene	12.853	95	141471	50.067	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
Data File : VN080238.D
Acq On : 22 Nov 2023 11:09
Operator : JC\MD
Sample : VN1122WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1122WBL01

Quant Time: Nov 23 00:15:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

