

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
 Data File : VN080242.D
 Acq On : 22 Nov 2023 12:54
 Operator : JC\MD
 Sample : 05516-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11930

Quant Time: Nov 23 00:18:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	148292	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	278359	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	322624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	138457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	141057	57.609	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.220%			
35) Dibromofluoromethane	8.171	113	109354	53.242	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.480%			
50) Toluene-d8	10.571	98	397601	53.701	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.400%			
62) 4-Bromofluorobenzene	12.853	95	154055	55.086	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 110.180%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	31054	37.760	ug/l	96
20) Methylene Chloride	5.289	84	38394	22.293	ug/l	95
43) Isopropyl Acetate	8.694	43	70970	17.681	ug/l #	87
95) Naphthalene	15.641	128	8334	1.057	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
Data File : VN080242.D
Acq On : 22 Nov 2023 12:54
Operator : JC\MD
Sample : 05516-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11930

Quant Time: Nov 23 00:18:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

