

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
 Data File : VN080243.D
 Acq On : 22 Nov 2023 13:18
 Operator : JC\MD
 Sample : 05516-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ETGI-343

Quant Time: Nov 23 00:19:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	149711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	292188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	328677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	144864	58.603	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	117.200%	
35) Dibromofluoromethane	8.171	113	110705	51.348	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.700%	
50) Toluene-d8	10.571	98	413579	53.215	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	106.440%	
62) 4-Bromofluorobenzene	12.853	95	156647	53.362	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	106.720%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	30734	37.017	ug/l #	85
20) Methylene Chloride	5.271	84	39686	22.825	ug/l	97
43) Isopropyl Acetate	8.688	43	90399	21.456	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
Data File : VN080243.D
Acq On : 22 Nov 2023 13:18
Operator : JC\MD
Sample : 05516-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ETGI-343

Quant Time: Nov 23 00:19:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

