

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
 Data File : VN080246.D
 Acq On : 22 Nov 2023 14:30
 Operator : JC\MD
 Sample : 05518-04 20X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1109

Quant Time: Nov 23 00:21:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	159994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	291639	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	315428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	149636	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	146376	55.409	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.820%			
35) Dibromofluoromethane	8.171	113	111893	51.997	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.000%			
50) Toluene-d8	10.570	98	389988	50.274	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.540%			
62) 4-Bromofluorobenzene	12.853	95	160948	54.931	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 109.860%			
Target Compounds			Qvalue			
39) Methylcyclohexane	9.600	83	3333	1.110	ug/l #	87
52) Toluene	10.635	92	6974	1.259	ug/l	97
68) m/p-Xylenes	12.070	106	12519	2.700	ug/l	83
69) o-Xylene	12.400	106	4599	1.052	ug/l	80
84) 1,2,4-Trimethylbenzene	13.488	105	14383	1.443	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112223\
Data File : VN080246.D
Acq On : 22 Nov 2023 14:30
Operator : JC\MD
Sample : 05518-04 20X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1109

Quant Time: Nov 23 00:21:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

