

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
 Data File : VN085012.D
 Acq On : 22 Nov 2024 16:07
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Nov 23 02:57:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	180278	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	293712	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	254182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	128770	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	129855	50.568	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.140%	
35) Dibromofluoromethane	8.159	113	102148	51.572	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.140%	
50) Toluene-d8	10.559	98	379679	53.488	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.980%	
62) 4-Bromofluorobenzene	12.847	95	141423	53.276	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.560%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.118	85	95019	46.612	ug/l	99
3) Chloromethane	2.359	50	99689	42.164	ug/l	98
4) Vinyl Chloride	2.512	62	97145	45.721	ug/l	95
5) Bromomethane	2.900	94	53525	44.945	ug/l	91
6) Chloroethane	3.077	64	59473	41.527	ug/l	94
7) Trichlorofluoromethane	3.477	101	165959	45.916	ug/l	94
8) Diethyl Ether	3.953	74	58978	47.942	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.359	101	93701	45.709	ug/l	99
10) Methyl Iodide	4.577	142	127705	47.063	ug/l	97
11) Tert butyl alcohol	5.536	59	74416	216.857	ug/l	97
12) 1,1-Dichloroethene	4.330	96	89314	46.677	ug/l	97
13) Acrolein	4.171	56	58781	256.710	ug/l	99
14) Allyl chloride	5.012	41	152182	47.178	ug/l	99
15) Acrylonitrile	5.712	53	225423	232.029	ug/l	99
16) Acetone	4.424	43	188963	252.665	ug/l	99
17) Carbon Disulfide	4.700	76	244899	43.362	ug/l #	94
18) Methyl Acetate	5.018	43	111346	46.500	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	303857	48.602	ug/l	98
20) Methylene Chloride	5.271	84	98339	43.090	ug/l	96
21) trans-1,2-Dichloroethene	5.777	96	92234	45.805	ug/l #	89
22) Diisopropyl ether	6.665	45	306957	47.281	ug/l	98
23) Vinyl Acetate	6.594	43	1108786	247.560	ug/l	100
24) 1,1-Dichloroethane	6.559	63	177343	45.647	ug/l	99
25) 2-Butanone	7.477	43	294434	240.802	ug/l	100
26) 2,2-Dichloropropane	7.488	77	168144	46.349	ug/l	98
27) cis-1,2-Dichloroethene	7.477	96	110716	45.609	ug/l	97
28) Bromochloromethane	7.812	49	67335	45.913	ug/l	97
29) Tetrahydrofuran	7.835	42	183433	241.241	ug/l	100
30) Chloroform	7.959	83	188550	45.623	ug/l	95
31) Cyclohexane	8.253	56	157639	46.084	ug/l	97
32) 1,1,1-Trichloroethane	8.159	97	170723	45.714	ug/l	99
36) 1,1-Dichloropropene	8.365	75	129026	48.093	ug/l	99
37) Ethyl Acetate	7.553	43	113781	46.651	ug/l	100
38) Carbon Tetrachloride	8.359	117	151265	48.846	ug/l	98
39) Methylcyclohexane	9.594	83	145286	52.114	ug/l	98
40) Benzene	8.600	78	404536	47.675	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
 Data File : VN085012.D
 Acq On : 22 Nov 2024 16:07
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Nov 23 02:57:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	66590	49.311	ug/l	98
42) 1,2-Dichloroethane	8.665	62	136508	47.025	ug/l	100
43) Isopropyl Acetate	8.688	43	210623	41.574	ug/l	99
44) Trichloroethene	9.347	130	93086	47.205	ug/l	98
45) 1,2-Dichloropropane	9.618	63	98380	47.073	ug/l	95
46) Dibromomethane	9.706	93	69575	47.909	ug/l	99
47) Bromodichloromethane	9.882	83	146603	47.631	ug/l	97
48) Methyl methacrylate	9.676	41	95959	50.900	ug/l	97
49) 1,4-Dioxane	9.694	88	31078	950.585	ug/l #	87
51) 4-Methyl-2-Pentanone	10.441	43	582480	251.720	ug/l	100
52) Toluene	10.629	92	249568	49.682	ug/l	99
53) t-1,3-Dichloropropene	10.829	75	150836	48.881	ug/l	97
54) cis-1,3-Dichloropropene	10.306	75	164009	50.100	ug/l	97
55) 1,1,2-Trichloroethane	11.012	97	92219	49.287	ug/l	99
56) Ethyl methacrylate	10.871	69	149537	54.694	ug/l	98
57) 1,3-Dichloropropane	11.159	76	158000	47.980	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	315602	235.323	ug/l	99
59) 2-Hexanone	11.194	43	444749	256.983	ug/l	100
60) Dibromochloromethane	11.359	129	108961	47.650	ug/l	99
61) 1,2-Dibromoethane	11.465	107	94657	48.883	ug/l	96
64) Tetrachloroethene	11.100	164	82753	48.316	ug/l	97
65) Chlorobenzene	11.888	112	264798	45.936	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	95374	46.991	ug/l	98
67) Ethyl Benzene	11.959	91	477400	51.046	ug/l	96
68) m/p-Xylenes	12.065	106	363006	101.951	ug/l	99
69) o-Xylene	12.394	106	173028	51.086	ug/l	98
70) Styrene	12.406	104	296759	52.687	ug/l	99
71) Bromoform	12.576	173	73244	48.617	ug/l #	98
73) Isopropylbenzene	12.694	105	445261	49.014	ug/l	100
74) N-amyl acetate	12.494	43	184990	49.418	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	132755	42.796	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	172673	73.722	ug/l	48
77) Bromobenzene	12.976	156	108566	44.410	ug/l	99
78) n-propylbenzene	13.035	91	525567	50.160	ug/l	98
79) 2-Chlorotoluene	13.123	91	324611	46.409	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	382819	50.220	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	49617	46.764	ug/l	98
82) 4-Chlorotoluene	13.217	91	331271	45.496	ug/l	99
83) tert-Butylbenzene	13.435	119	320979	48.939	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	392031	49.962	ug/l	99
85) sec-Butylbenzene	13.612	105	444471	50.166	ug/l	99
86) p-Isopropyltoluene	13.729	119	378270	51.719	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	204731	42.459	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	202930	41.240	ug/l	97
89) n-Butylbenzene	14.053	91	325752	47.892	ug/l	100
90) Hexachloroethane	14.335	117	72271	45.332	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	200864	43.472	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	26503	44.183	ug/l	99
93) 1,2,4-Trichlorobenzene	15.394	180	107707	42.908	ug/l	99
94) Hexachlorobutadiene	15.500	225	49330	42.581	ug/l	99
95) Naphthalene	15.641	128	329352	46.067	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	102829	42.932	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
Data File : VN085012.D
Acq On : 22 Nov 2024 16:07
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Nov 23 02:57:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Nov 23 02:57:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:168 Resp: 180278
Ion Ratio Lower Upper
168 100
99 62.4 51.2 76.8



#2
Dichlorodifluoromethane
Concen: 46.612 ug/l
RT: 2.118 min Scan# 28
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 85 Resp: 95019
Ion Ratio Lower Upper
85 100
87 31.9 15.7 47.1









#3
Chloromethane
Concen: 42.164 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 99689
Ion Ratio Lower Upper
50 100
52 33.8 27.8 41.8



#4
Vinyl Chloride
Concen: 45.721 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 62 Resp: 97145
Ion Ratio Lower Upper
62 100
64 33.2 24.6 36.8









#5
Bromomethane
Concen: 44.945 ug/l
RT: 2.900 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 53525
Ion Ratio Lower Upper
94 100
96 89.2 78.7 118.1



#6
Chloroethane
Concen: 41.527 ug/l
RT: 3.077 min Scan# 191
Delta R.T. 0.012 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 64 Resp: 59473
Ion Ratio Lower Upper
64 100
66 34.0 24.4 36.6









#7
 Trichlorofluoromethane
 Concen: 45.916 ug/l
 RT: 3.477 min Scan# 20
 Delta R.T. 0.017 min
 Lab File: VN085012.D
 Acq: 22 Nov 2024 16:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:101 Resp: 165959
 Ion Ratio Lower Upper
 101 100
 103 61.3 52.6 78.8



#8
 Diethyl Ether
 Concen: 47.942 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. 0.006 min
 Lab File: VN085012.D
 Acq: 22 Nov 2024 16:07

Tgt Ion: 74 Resp: 58978
 Ion Ratio Lower Upper
 74 100
 45 95.6 46.7 140.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.709 ug/l
RT: 4.359 min Scan# 409
Delta R.T. 0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 93701
Ion Ratio Lower Upper
101 100
85 45.7 36.9 55.3
151 75.2 60.8 91.2



#10
Methyl Iodide
Concen: 47.063 ug/l
RT: 4.577 min Scan# 446
Delta R.T. 0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07


Tgt Ion:142 Resp: 127705
Ion Ratio Lower Upper
142 100
127 47.6 36.6 54.8
141 13.3 12.2 18.4








#11
Tert butyl alcohol
Concen: 216.857 ug/l
RT: 5.536 min Scan# 609
Delta R.T. -0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 59 Resp: 74416
Ion Ratio Lower Upper
59 100
57 10.1 7.4 11.0



#12
1,1-Dichloroethene
Concen: 46.677 ug/l
RT: 4.330 min Scan# 404
Delta R.T. 0.012 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 96 Resp: 89314
Ion Ratio Lower Upper
96 100
61 164.2 127.8 191.8
98 63.7 49.2 73.8





#13

Acrolein

Concen: 256.710 ug/l

RT: 4.171 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 56 Resp: 58781

Ion Ratio Lower Upper

56 100

55 69.2 54.5 81.7



#14

Allyl chloride

Concen: 47.178 ug/l

RT: 5.012 min Scan# 520

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07




Tgt Ion: 41 Resp: 152182


Ion Ratio Lower Upper

41 100

39 66.6 53.8 80.8

76 35.9 29.3 43.9









#15
Acrylonitrile
Concen: 232.029 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 53 Resp: 225423
Ion Ratio Lower Upper
53 100
52 81.5 65.5 98.3
51 34.6 28.4 42.6



#16
Acetone
Concen: 252.665 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 43 Resp: 188963
Ion Ratio Lower Upper
43 100
58 31.1 25.4 38.2









#17
Carbon Disulfide
Concen: 43.362 ug/l
RT: 4.700 min Scan# 467
Delta R.T. 0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 244899
Ion Ratio Lower Upper
76 100
78 10.1 6.5 9.7#



#18
Methyl Acetate
Concen: 46.500 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 43 Resp: 111346
Ion Ratio Lower Upper
43 100
74 24.4 20.1 30.1









#19
Methyl tert-butyl Ether
Concen: 48.602 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 73 Resp: 303857
Ion Ratio Lower Upper
73 100
57 21.9 17.0 25.4



#20
Methylene Chloride
Concen: 43.090 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 84 Resp: 98339
Ion Ratio Lower Upper
84 100
49 110.0 92.7 139.1
51 36.7 29.4 44.2
86 62.6 51.9 77.9





#21

trans-1,2-Dichloroethene

Concen: 45.805 ug/l

RT: 5.777 min Scan# 649

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 96 Resp: 92234

Ion Ratio Lower Upper

96 100

61 136.1 100.7 151.1

98 68.7 45.8 68.6#



#22

Diisopropyl ether

Concen: 47.281 ug/l

RT: 6.665 min Scan# 801

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Tgt Ion: 45 Resp: 306957




Ion Ratio Lower Upper


45 100

43 55.0 42.7 64.1

87 28.5 22.6 34.0

59 11.3 9.9 14.9









#23
 Vinyl Acetate
 Concen: 247.560 ug/l
 RT: 6.594 min Scan# 789
 Delta R.T. -0.000 min
 Lab File: VN085012.D
 Acq: 22 Nov 2024 16:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 1108786
 Ion Ratio Lower Upper
 43 100
 86 11.0 8.9 13.3



#24
 1,1-Dichloroethane
 Concen: 45.647 ug/l
 RT: 6.559 min Scan# 783
 Delta R.T. 0.006 min
 Lab File: VN085012.D
 Acq: 22 Nov 2024 16:07

Tgt Ion: 63 Resp: 177343
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.3 9.8
 100 4.8 2.2 6.6









#25
2-Butanone
Concen: 240.802 ug/l
RT: 7.477 min Scan# 940
Delta R.T. -0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 294434
Ion Ratio Lower Upper
43 100
72 26.5 21.4 32.0



#26
2,2-Dichloropropane
Concen: 46.349 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07


Tgt Ion: 77 Resp: 168144
Ion Ratio Lower Upper
77 100
97 22.5 10.9 32.7








#27
 cis-1,2-Dichloroethene
 Concen: 45.609 ug/l
 RT: 7.477 min Scan# 939
 Delta R.T. -0.000 min
 Lab File: VN085012.D
 Acq: 22 Nov 2024 16:07

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 96 Resp: 110716
 Ion Ratio Lower Upper
 96 100
 61 149.3 0.0 291.6
 98 66.6 0.0 129.8



#28
 Bromochloromethane
 Concen: 45.913 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085012.D
 Acq: 22 Nov 2024 16:07


Tgt Ion: 49 Resp: 67335
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.0
 130 78.0 60.6 90.8








#29
Tetrahydrofuran
Concen: 241.241 ug/l
RT: 7.835 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 42 Resp: 183433
Ion Ratio Lower Upper
42 100
72 46.4 37.0 55.4
71 43.6 35.0 52.4



#30
Chloroform
Concen: 45.623 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07


Tgt Ion: 83 Resp: 188550
Ion Ratio Lower Upper
83 100
85 61.8 52.6 78.8








#31
Cyclohexane
Concen: 46.084 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 157639
Ion Ratio Lower Upper
56 100
69 30.0 24.6 36.8
84 85.3 71.4 107.2



#32
1,1,1-Trichloroethane
Concen: 45.714 ug/l
RT: 8.159 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 97 Resp: 170723
Ion Ratio Lower Upper
97 100
99 64.1 51.0 76.4
61 45.6 36.2 54.2









#33
1,2-Dichloroethane-d4
Concen: 50.568 ug/l
RT: 8.571 min Scan# 1125
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 65 Resp: 129855
Ion Ratio Lower Upper
65 100
67 51.6 0.0 105.0



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion:114 Resp: 293712
Ion Ratio Lower Upper
114 100
63 20.7 0.0 41.0
88 16.1 0.0 34.8





#35

Dibromofluoromethane

Concen: 51.572 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:113 Resp: 102148

Ion Ratio Lower Upper

113 100

111 105.3 82.0 123.0

192 17.4 14.8 22.2



#36

1,1-Dichloropropene

Concen: 48.093 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07




Tgt Ion: 75 Resp: 129026


Ion Ratio Lower Upper

75 100

110 35.2 17.2 51.6

77 30.7 24.2 36.4





#37

Ethyl Acetate

Concen: 46.651 ug/l

RT: 7.553 min Scan# 91

Delta R.T. -0.000 min MSVOA_N

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 43 Resp: 113781

Ion Ratio Lower Upper

43 100

61 14.9 11.7 17.5

70 11.5 9.3 13.9



#38

Carbon Tetrachloride

Concen: 48.846 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07




Tgt Ion: 117 Resp: 151265


Ion Ratio Lower Upper

117 100

119 95.5 78.6 117.8

121 30.3 24.2 36.2





#39

Methylcyclohexane

Concen: 52.114 ug/l

RT: 9.594 min Scan# 1129

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 145286

Ion Ratio Lower Upper

83 100

55 79.2 61.8 92.6

98 48.1 37.0 55.6



#40

Benzene

Concen: 47.675 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN085012.D




Acq: 22 Nov 2024 16:07


Tgt Ion: 78 Resp: 404536

Ion Ratio Lower Upper

78 100


77 23.6 18.5 27.7








#41
Methacrylonitrile
Concen: 49.311 ug/l
RT: 7.771 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:	41	Resp:	66590
Ion	Ratio	Lower	Upper
41	100		
39	56.6	42.6	63.8
67	75.5	59.9	89.9
52	30.9	24.8	37.2



#42
1,2-Dichloroethane
Concen: 47.025 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion:	62	Resp:	136508
Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	18.2





#43

Isopropyl Acetate

Concen: 41.574 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 43 Resp: 210623

Ion Ratio Lower Upper

43 100

61 26.7 20.6 30.8

87 13.5 11.0 16.4



#44

Trichloroethene

Concen: 47.205 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN085012.D




Acq: 22 Nov 2024 16:07


Tgt Ion:130 Resp: 93086

Ion Ratio Lower Upper

130 100

95 108.7 0.0 213.4









#45
1,2-Dichloropropane
Concen: 47.073 ug/l
RT: 9.618 min Scan# 1303
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 63 Resp: 98380
Ion Ratio Lower Upper
63 100
65 32.5 24.0 36.0



#46
Dibromomethane
Concen: 47.909 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07


Tgt Ion: 93 Resp: 69575
Ion Ratio Lower Upper
93 100
95 82.9 66.2 99.4
174 91.9 74.9 112.3








#47
Bromodichloromethane
Concen: 47.631 ug/l
RT: 9.882 min Scan# 1348
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 83 Resp: 146603
Ion Ratio Lower Upper
83 100
85 66.2 50.5 75.7
127 8.9 7.1 10.7



#48
Methyl methacrylate
Concen: 50.900 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 41 Resp: 95959
Ion Ratio Lower Upper
41 100
69 86.9 73.2 109.8
39 57.0 46.2 69.4





#49

1,4-Dioxane

Concen: 950.585 ug/l

RT: 9.694 min Scan# 14

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 88 Resp: 31078

Ion Ratio Lower Upper

88 100

43 0.0 16.7 25.1#

58 73.0 56.5 84.7



#50

Toluene-d8

Concen: 53.488 ug/l

RT: 10.559 min Scan# 1463

Delta R.T. -0.006 min

Lab File: VN085012.D




Acq: 22 Nov 2024 16:07


Tgt Ion: 98 Resp: 379679

Ion Ratio Lower Upper

98 100

100 66.2 52.5 78.7





#51

4-Methyl-2-Pentanone

Concen: 251.720 ug/l

RT: 10.441 min Scan# 1443

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 43 Resp: 582480

Ion Ratio Lower Upper

43 100

58 40.4 32.1 48.1



#52

Toluene

Concen: 49.682 ug/l

RT: 10.629 min Scan# 1475

Delta R.T. 0.006 min

Lab File: VN085012.D




Acq: 22 Nov 2024 16:07


Tgt Ion: 92 Resp: 249568

Ion Ratio Lower Upper

92 100

91 171.9 136.7 205.1









#53
t-1,3-Dichloropropene
Concen: 48.881 ug/l
RT: 10.829 min Scan# 1510
Delta R.T. -0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 150836
Ion Ratio Lower Upper
75 100
77 29.7 25.0 37.4



#54
cis-1,3-Dichloropropene
Concen: 50.100 ug/l
RT: 10.306 min Scan# 1420
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07


Tgt Ion: 75 Resp: 164009
Ion Ratio Lower Upper
75 100
77 29.6 25.0 37.4
39 48.9 37.6 56.4








#55
1,1,2-Trichloroethane
Concen: 49.287 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :




Tgt Ion: 97 Resp: 92219

Ion	Ratio	Lower	Upper
97	100		
83	83.5	67.4	101.2
85	53.8	43.1	64.7
99	63.9	50.9	76.3






#56
Ethyl methacrylate
Concen: 54.694 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07



Tgt Ion: 69 Resp: 149537

Ion	Ratio	Lower	Upper
69	100		
41	61.1	50.0	75.0
39	34.6	27.0	40.6









#57
1,3-Dichloropropane
Concen: 47.980 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 158000
Ion Ratio Lower Upper
76 100
78 32.0 26.1 39.1



#58
2-Chloroethyl Vinyl ether
Concen: 235.323 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 63 Resp: 315602
Ion Ratio Lower Upper
63 100
106 29.1 23.0 34.6









#59
2-Hexanone
Concen: 256.983 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 444749
Ion Ratio Lower Upper
43 100
58 56.9 28.3 85.0



#60
Dibromochloromethane
Concen: 47.650 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion:129 Resp: 108961
Ion Ratio Lower Upper
129 100
127 79.0 39.2 117.6









#61
1,2-Dibromoethane
Concen: 48.883 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 94657
Ion Ratio Lower Upper
107 100
109 90.8 76.1 114.1



#62
4-Bromofluorobenzene
Concen: 53.276 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 95 Resp: 141423
Ion Ratio Lower Upper
95 100
174 73.4 0.0 146.6
176 69.8 0.0 141.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:117 Resp: 254182

Ion Ratio Lower Upper

117 100

82 55.3 46.0 69.0

119 30.8 25.6 38.4



#64

Tetrachloroethene

Concen: 48.316 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Tgt Ion:164 Resp: 82753




Ion Ratio Lower Upper


164 100

166 122.9 100.6 151.0

129 94.2 77.7 116.5


131 92.5 77.2 115.8








#65
Chlorobenzene
Concen: 45.936 ug/l
RT: 11.888 min Scan# 112
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:112 Resp: 264798
Ion Ratio Lower Upper
112 100
114 32.1 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 46.991 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion:131 Resp: 95374
Ion Ratio Lower Upper
131 100
133 97.7 47.8 143.4
119 65.5 32.1 96.3





#67

Ethyl Benzene

Concen: 51.046 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 91 Resp: 477400

Ion Ratio Lower Upper

91 100

106 28.8 24.7 37.1



#68

m/p-Xylenes

Concen: 101.951 ug/l

RT: 12.065 min Scan# 1719

Delta R.T. -0.006 min

Lab File: VN085012.D




Acq: 22 Nov 2024 16:07


Tgt Ion:106 Resp: 363006

Ion Ratio Lower Upper

106 100

91 209.0 168.0 252.0





#69

o-Xylene

Concen: 51.086 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:106 Resp: 173028

Ion Ratio Lower Upper

106 100

91 223.8 110.7 331.9



#70

Styrene

Concen: 52.687 ug/l

RT: 12.406 min Scan# 1777

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07




Tgt Ion:104 Resp: 296759


Ion Ratio Lower Upper

104 100

78 52.3 42.3 63.5

103 55.0 44.2 66.2









#71
Bromoform
Concen: 48.617 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 73244
Ion Ratio Lower Upper
173 100
175 49.0 23.8 71.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 1212
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion:152 Resp: 128770
Ion Ratio Lower Upper
152 100
115 63.6 31.7 95.1
150 173.6 0.0 351.4









#73
Isopropylbenzene
Concen: 49.014 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 445261
Ion Ratio Lower Upper
105 100
120 26.0 13.0 39.0



#74
N-ethyl acetate
Concen: 49.418 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 43 Resp: 184990
Ion Ratio Lower Upper
43 100
70 44.4 34.6 51.8
55 25.8 20.5 30.7
61 25.3 19.6 29.4









#75
1,1,2,2-Tetrachloroethane
Concen: 42.796 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 132755
Ion Ratio Lower Upper
83 100
131 10.6 5.1 15.4
85 64.0 32.1 96.3



#76
1,2,3-Trichloropropane
Concen: 73.722 ug/l
RT: 12.988 min Scan# 1876
Delta R.T. -0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07


Tgt Ion: 75 Resp: 172673
Ion Ratio Lower Upper
75 100
77 135.8 109.1 327.4








#77
Bromobenzene
Concen: 44.410 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:156 Resp: 108566
Ion Ratio Lower Upper
156 100
77 216.1 107.2 321.6
158 97.6 48.5 145.6



#78
n-propylbenzene
Concen: 50.160 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 91 Resp: 525567
Ion Ratio Lower Upper
91 100
120 21.8 11.3 33.9









#79
2-Chlorotoluene
Concen: 46.409 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 324611
Ion Ratio Lower Upper
91 100
126 32.2 16.1 48.3



#80
1,3,5-Trimethylbenzene
Concen: 50.220 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion:105 Resp: 382819
Ion Ratio Lower Upper
105 100
120 47.0 23.6 71.0









#81
trans-1,4-Dichloro-2-butene
Concen: 46.764 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 49617
Ion Ratio Lower Upper
75 100
53 96.8 78.2 117.4
89 44.2 37.9 56.9



#82
4-Chlorotoluene
Concen: 45.496 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 91 Resp: 331271
Ion Ratio Lower Upper
91 100
126 32.5 16.1 48.3





#83

tert-Butylbenzene

Concen: 48.939 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:119 Resp: 320979

Ion Ratio Lower Upper

119 100

91 70.2 35.1 105.5

134 27.2 12.1 36.3



#84

1,2,4-Trimethylbenzene

Concen: 49.962 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.006 min

Lab File: VN085012.D




Acq: 22 Nov 2024 16:07


Tgt Ion:105 Resp: 392031

Ion Ratio Lower Upper

105 100

120 43.1 21.8 65.4





#85

sec-Butylbenzene

Concen: 50.166 ug/l

RT: 13.612 min Scan# 1982

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 444471

Ion Ratio Lower Upper

105 100

134 19.9 9.8 29.3



#86

p-Isopropyltoluene

Concen: 51.719 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07




Tgt Ion:119 Resp: 378270


Ion Ratio Lower Upper

119 100

134 25.6 13.0 39.0

91 25.6 12.8 38.4





#87

1,3-Dichlorobenzene

Concen: 42.459 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:146 Resp: 204731

Ion Ratio Lower Upper

146 100

111 42.9 21.6 65.0

148 64.4 32.2 96.6



#88

1,4-Dichlorobenzene

Concen: 41.240 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07




Tgt Ion:146 Resp: 202930


Ion Ratio Lower Upper

146 100

111 43.0 20.9 62.8


148 65.1 31.3 93.9








#89
n-Butylbenzene
Concen: 47.892 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :



Tgt Ion: 91 Resp: 325752




Ion	Ratio	Lower	Upper
91	100		
92	53.2	26.8	80.3
134	25.4	12.8	38.4




#90
Hexachloroethane
Concen: 45.332 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.006 min
Lab File: VN085012.D
Acq: 22 Nov 2024 16:07

Tgt Ion: 117 Resp: 72271

Ion	Ratio	Lower	Upper
117	100		
201	68.7	35.2	105.6





#91

1,2-Dichlorobenzene

Concen: 43.472 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:146 Resp: 200864

Ion Ratio Lower Upper

146 100

111 44.4 21.9 65.8

148 63.7 31.1 93.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 44.183 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07




Tgt Ion: 75 Resp: 26503


Ion Ratio Lower Upper

75 100

155 76.6 38.9 116.7

157 95.3 48.3 144.9





#93

1,2,4-Trichlorobenzene

Concen: 42.908 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:180 Resp: 107707

Ion Ratio Lower Upper

180 100

182 94.4 46.9 140.8

145 34.3 16.3 48.8



#94

Hexachlorobutadiene

Concen: 42.581 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07




Tgt Ion:225 Resp: 49330


Ion Ratio Lower Upper

225 100

223 61.4 30.9 92.8

227 61.9 31.5 94.5





#95

Naphthalene

Concen: 46.067 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:128 Resp: 329352

Ion Ratio Lower Upper

128 100

127 12.8 10.6 16.0

129 10.7 8.6 12.8



#96

1,2,3-Trichlorobenzene

Concen: 42.932 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN085012.D

Acq: 22 Nov 2024 16:07

Tgt Ion:180 Resp: 102829

Ion Ratio Lower Upper

180 100

182 95.9 46.4 139.2

145 36.1 17.3 52.0

