

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
 Data File : VN085011.D
 Acq On : 22 Nov 2024 14:09
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Nov 23 02:56:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156982	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	273594	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	236620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100776	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113482	50.750	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	101.500%	
35) Dibromofluoromethane	8.159	113	90800	49.213	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.420%	
50) Toluene-d8	10.565	98	312193	47.215	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.440%	
62) 4-Bromofluorobenzene	12.847	95	104339	42.196	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	84.400%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
Data File : VN085011.D
Acq On : 22 Nov 2024 14:09
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Nov 23 02:56:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

