

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
 Data File : VN085016.D
 Acq On : 22 Nov 2024 19:17
 Operator : JC\MD
 Sample : P4845-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-13-20241112

Quant Time: Nov 23 03:25:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	152210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	264627	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	222942	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	99476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109942	50.708	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	101.420%		
35) Dibromofluoromethane	8.159	113	86430	48.432	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.860%		
50) Toluene-d8	10.565	98	305909	47.832	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.660%		
62) 4-Bromofluorobenzene	12.847	95	109809	45.913	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	91.820%		
Target Compounds						Qvalue
44) Trichloroethene	9.347	130	205953	115.921	ug/l	94
64) Tetrachloroethene	11.094	164	1116	0.743	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
Data File : VN085016.D
Acq On : 22 Nov 2024 19:17
Operator : JC\MD
Sample : P4845-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-13-20241112

Quant Time: Nov 23 03:25:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

