

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
 Data File : VN085018.D
 Acq On : 22 Nov 2024 20:05
 Operator : JC\MD
 Sample : P4845-23
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-18-20241112

Quant Time: Nov 23 03:26:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/25/2024
 Supervised By : Mahesh Dadoda 11/25/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	169757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	301590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	117323	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	124950	51.674	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.340%			
35) Dibromofluoromethane	8.159	113	99028	48.691	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.380%			
50) Toluene-d8	10.565	98	346771	47.576	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.160%			
62) 4-Bromofluorobenzene	12.847	95	127974	46.950	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.900%			
Target Compounds						
16) Acetone	4.430	43	2023m	2.873	ug/l	
44) Trichloroethene	9.347	130	88308	43.613	ug/l	90
64) Tetrachloroethene	11.100	164	2087	1.169	ug/l #	87
93) 1,2,4-Trichlorobenzene	15.400	180	581	0.254	ug/l	92
95) Naphthalene	15.647	128	1736	0.267	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
Data File : VN085018.D
Acq On : 22 Nov 2024 20:05
Operator : JC\MD
Sample : P4845-23
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-18-20241112

Quant Time: Nov 23 03:26:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/25/2024
Supervised By :Mahesh Dadoda 11/25/2024

