

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
 Data File : VN085019.D
 Acq On : 22 Nov 2024 20:29
 Operator : JC\MD
 Sample : P4845-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-20-20241112

Quant Time: Nov 23 03:27:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/25/2024
 Supervised By : Mahesh Dadoda 11/25/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	172730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	267127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127888	51.978	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.960%			
35) Dibromofluoromethane	8.159	113	100722	49.540	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.080%			
50) Toluene-d8	10.565	98	349119	47.914	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.820%			
62) 4-Bromofluorobenzene	12.847	95	135797	49.836	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 99.680%			
Target Compounds						
						Qvalue
16) Acetone	4.442	43	3085m	4.305	ug/l	
44) Trichloroethene	9.347	130	67963	33.576	ug/l	98
64) Tetrachloroethene	11.100	164	2648	1.471	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112224\
Data File : VN085019.D
Acq On : 22 Nov 2024 20:29
Operator : JC\MD
Sample : P4845-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-20-20241112

Quant Time: Nov 23 03:27:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/25/2024
Supervised By :Mahesh Dadoda 11/25/2024

