

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112320\
 Data File : VN064784.D
 Acq On : 23 Nov 2020 15:20
 Operator : JC/MD
 Sample : VN1123WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1123WBL01

Quant Time: Nov 24 03:41:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	152878	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	270448	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	279648	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	127208	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96629	54.37	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	108.74%		
35) Dibromofluoromethane	7.55	113	86209	53.27	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	106.54%		
50) Toluene-d8	10.06	98	339742	51.38	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.76%		
62) 4-Bromofluorobenzene	12.38	95	131761	55.68	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	111.36%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112320\
Data File : VN064784.D
Acq On : 23 Nov 2020 15:20
Operator : JC/MD
Sample : VN1123WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1123WBL01

Quant Time: Nov 24 03:41:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

