

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112320\
 Data File : VN064789.D
 Acq On : 23 Nov 2020 17:34
 Operator : JC/MD
 Sample : L4829-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 24 03:57:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	152367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	277291	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	287728	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	131659	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	98763	55.76	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.52%	
35) Dibromofluoromethane		7.55	113	87391	52.67	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.34%	
50) Toluene-d8		10.06	98	344603	50.83	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.66%	
62) 4-Bromofluorobenzene		12.38	95	134641	55.49	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.98%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112320\
Data File : VN064789.D
Acq On : 23 Nov 2020 17:34
Operator : JC/MD
Sample : L4829-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 24 03:57:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

