

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112320\
 Data File : VN064795.D
 Acq On : 24 Nov 2020 1:09
 Operator : JC/MD
 Sample : L4829-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17

Quant Time: Nov 24 04:49:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	156712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.54	114	275241	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	244733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.31	152	86271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	107291	58.89	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	117.78%	
35) Dibromofluoromethane	7.54	113	94393	57.31	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	114.62%	
50) Toluene-d8	10.05	98	337962	50.22	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	100.44%	
62) 4-Bromofluorobenzene	12.37	95	101436	42.12	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	84.24%	
Target Compounds						
16) Acetone	3.78	43	4221	7.048	ug/l	98
30) Chloroform	7.32	83	6851	2.216	ug/l	93
64) Tetrachloroethene	10.59	164	2339	0.982	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112320\
Data File : VN064795.D
Acq On : 24 Nov 2020 1:09
Operator : JC/MD
Sample : L4829-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-17

Quant Time: Nov 24 04:49:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

