

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075386.D
 Acq On : 23 Nov 2022 10:53
 Operator : JC\MD
 Sample : VN1123MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1123MBL01

Quant Time: Nov 23 11:10:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	163819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	281092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	246611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	105477	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	79130	89.986	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	179.980%#
35) Dibromofluoromethane	8.171	113	69335	62.791	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	125.580%#
50) Toluene-d8	10.570	98	252363	101.472	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	202.940%#
62) 4-Bromofluorobenzene	12.847	95	82422	44.900	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075386.D
Acq On : 23 Nov 2022 10:53
Operator : JC\MD
Sample : VN1123MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1123MBL01

Quant Time: Nov 23 11:10:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

