

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075392.D
 Acq On : 23 Nov 2022 13:36
 Operator : JC\MD
 Sample : N5675-01 5X
 Misc : 1.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1673DL

Quant Time: Nov 23 13:56:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	127448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	215939	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	191671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	57830	84.532	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 169.060%#		
35) Dibromofluoromethane	8.171	113	70030	82.555	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 165.120%#		
50) Toluene-d8	10.571	98	157052	82.202	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 164.400%#		
62) 4-Bromofluorobenzene	12.847	95	116504	82.615	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 165.240%#		
Target Compounds						
						Qvalue
52) Toluene	10.629	92	85825	23.515	ug/l	99
67) Ethyl Benzene	11.965	91	1919	0.284	ug/l	96
68) m/p-Xylenes	12.065	106	2008	0.732	ug/l #	64
80) 1,3,5-Trimethylbenzene	13.171	105	1769	0.320	ug/l	77
84) 1,2,4-Trimethylbenzene	13.482	105	7578	1.365	ug/l	99
95) Naphthalene	15.641	128	4018	0.820	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075392.D
Acq On : 23 Nov 2022 13:36
Operator : JC\MD
Sample : N5675-01 5X
Misc : 1.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1673DL

Quant Time: Nov 23 13:56:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

