

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075393.D
 Acq On : 23 Nov 2022 13:59
 Operator : JC\MD
 Sample : N5692-03 5X
 Misc : 1.32g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HS-SPILL-DRUM-11-16

Quant Time: Nov 24 06:51:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	164104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	277484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	252439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125204	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	41468	47.075	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.160%
35) Dibromofluoromethane	8.171	113	47875	43.920	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.840%
50) Toluene-d8	10.565	98	107098	43.623	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.240%
62) 4-Bromofluorobenzene	12.847	95	80974	44.685	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075393.D
Acq On : 23 Nov 2022 13:59
Operator : JC\MD
Sample : N5692-03 5X
Misc : 1.32g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HS-SPILL-DRUM-11-16

Quant Time: Nov 24 06:51:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

