

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075400.D
 Acq On : 23 Nov 2022 16:47
 Operator : JC\MD
 Sample : PB149245ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149245ZHE#03

Quant Time: Nov 24 06:52:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	162979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	287328	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	266276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	46274	52.894	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.780%
35) Dibromofluoromethane	8.171	113	55849	49.480	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.960%
50) Toluene-d8	10.571	98	129698	51.018	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.040%
62) 4-Bromofluorobenzene	12.853	95	92906	49.513	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.020%
Target Compounds						
						Qvalue
8) Diethyl Ether	3.965	74	4014	4.102	ug/l	91
16) Acetone	4.447	43	83282	102.059	ug/l	96
20) Methylene Chloride	5.289	84	21996	12.329	ug/l #	82
37) Ethyl Acetate	7.571	43	22862	10.315	ug/l	98
43) Isopropyl Acetate	8.694	43	148497	41.741	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075400.D
Acq On : 23 Nov 2022 16:47
Operator : JC\MD
Sample : PB149245ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149245ZHE#03

Quant Time: Nov 24 06:52:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

