

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075402.D
 Acq On : 23 Nov 2022 17:35
 Operator : JC\MD
 Sample : PB149245ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149245ZHE#05

Quant Time: Nov 24 06:53:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	159434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	281775	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	255494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	43401	50.713	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.420%
35) Dibromofluoromethane	8.171	113	52635	47.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	10.571	98	122877	49.288	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.580%
62) 4-Bromofluorobenzene	12.853	95	86832	47.188	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.380%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.977	74	2850	2.977	ug/l	91
16) Acetone	4.442	43	59563	74.615	ug/l	94
18) Methyl Acetate	5.036	43	4734	1.085	ug/l #	86
20) Methylene Chloride	5.289	84	17228	9.610	ug/l	86
37) Ethyl Acetate	7.577	43	22165	10.198	ug/l	98
43) Isopropyl Acetate	8.694	43	182352	52.268	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075402.D
Acq On : 23 Nov 2022 17:35
Operator : JC\MD
Sample : PB149245ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149245ZHE#05

Quant Time: Nov 24 06:53:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

