

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075403.D
 Acq On : 23 Nov 2022 17:59
 Operator : JC\MD
 Sample : N5758-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-13

Quant Time: Nov 24 06:53:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	151596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	269965	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	243054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	106810	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	45101	55.424	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.840%
35) Dibromofluoromethane	8.171	113	53812	50.742	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.480%
50) Toluene-d8	10.571	98	123882	51.865	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.720%
62) 4-Bromofluorobenzene	12.847	95	88231	50.045	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.100%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	17773	23.416	ug/l	99
20) Methylene Chloride	5.289	84	8736	4.512	ug/l #	86
37) Ethyl Acetate	7.571	43	19211	9.225	ug/l	98
43) Isopropyl Acetate	8.694	43	174389	52.172	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075403.D
Acq On : 23 Nov 2022 17:59
Operator : JC\MD
Sample : N5758-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-13

Quant Time: Nov 24 06:53:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

