

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075404.D
 Acq On : 23 Nov 2022 18:23
 Operator : JC\MD
 Sample : N5762-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Nov 24 06:53:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	158230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	277642	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255942	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	43235	50.903	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.800%
35) Dibromofluoromethane	8.171	113	50829	46.604	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.200%
50) Toluene-d8	10.570	98	119459	48.630	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.260%
62) 4-Bromofluorobenzene	12.853	95	87241	48.116	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.240%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	19590	24.727	ug/l	98
18) Methyl Acetate	5.041	43	4897	1.188	ug/l	99
20) Methylene Chloride	5.283	84	8480	4.105	ug/l	86
37) Ethyl Acetate	7.571	43	19754	9.224	ug/l	98
43) Isopropyl Acetate	8.694	43	140102	40.755	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075404.D
Acq On : 23 Nov 2022 18:23
Operator : JC\MD
Sample : N5762-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Nov 24 06:53:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

