

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075406.D
 Acq On : 23 Nov 2022 19:10
 Operator : JC\MD
 Sample : N5765-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-L

Quant Time: Nov 24 06:53:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	158831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	277112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	255443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	45091	52.888	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.780%	
35) Dibromofluoromethane	8.171	113	55248	50.752	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.500%	
50) Toluene-d8	10.571	98	125239	51.080	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.160%	
62) 4-Bromofluorobenzene	12.853	95	91872	50.767	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.540%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	16988	21.362	ug/l	95
18) Methyl Acetate	5.059	43	4755	1.105	ug/l	94
20) Methylene Chloride	5.289	84	8605	4.164	ug/l #	82
37) Ethyl Acetate	7.577	43	19027	8.901	ug/l	100
43) Isopropyl Acetate	8.694	43	151289	44.094	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075406.D
Acq On : 23 Nov 2022 19:10
Operator : JC\MD
Sample : N5765-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-L

Quant Time: Nov 24 06:53:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

