

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
 Data File : VN075409.D
 Acq On : 23 Nov 2022 20:22
 Operator : JC\MD
 Sample : N5692-03
 Misc : 1.32g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HS-SPILL-DRUM-11-16

Quant Time: Nov 24 06:54:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	258620	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	223641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	102260	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	38170	46.261	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.520%
35) Dibromofluoromethane	8.171	113	44672	43.971	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.940%
50) Toluene-d8	10.570	98	103098	45.057	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.120%
62) 4-Bromofluorobenzene	12.853	95	72117	42.700	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.400%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.482	105	9202	1.478	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112322\
Data File : VN075409.D
Acq On : 23 Nov 2022 20:22
Operator : JC\MD
Sample : N5692-03
Misc : 1.32g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HS-SPILL-DRUM-11-16

Quant Time: Nov 24 06:54:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

