

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064800.D
 Acq On : 24 Nov 2020 10:52
 Operator : JC/MD
 Sample : VN1124WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1124WBL01

Quant Time: Nov 25 04:15:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	160151	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	284754	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	293941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136738	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	101731	54.64	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.28%	
35) Dibromofluoromethane		7.55	113	91160	53.50	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.00%	
50) Toluene-d8		10.06	98	356461	51.20	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.40%	
62) 4-Bromofluorobenzene		12.38	95	137305	55.11	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064800.D
Acq On : 24 Nov 2020 10:52
Operator : JC/MD
Sample : VN1124WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1124WBL01

Quant Time: Nov 25 04:15:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

