

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064809.D
 Acq On : 24 Nov 2020 14:39
 Operator : JC/MD
 Sample : L4829-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-8

Quant Time: Nov 25 04:58:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	174945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	311423	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	317444	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	147933	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	105136	51.69	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.38%	
35) Dibromofluoromethane		7.55	113	96412	51.74	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.48%	
50) Toluene-d8		10.06	98	387742	50.93	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.86%	
62) 4-Bromofluorobenzene		12.38	95	153063	56.17	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	1530	2.288	ug/l #	80
30) Chloroform	7.33	83	64078	18.565	ug/l	98
64) Tetrachloroethene	10.60	164	1919	0.621	ug/l #	83
84) 1,2,4-Trimethylbenzene	13.01	105	2041	0.236	ug/l	97
95) Naphthalene	15.10	128	7268	0.810	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064809.D
Acq On : 24 Nov 2020 14:39
Operator : JC/MD
Sample : L4829-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-8

Quant Time: Nov 25 04:58:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

