

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064812.D
 Acq On : 24 Nov 2020 15:51
 Operator : JC/MD
 Sample : L4829-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11

Manual Integrations
 APPROVED

MMDadoda
 11/25/2020 3:57:24 PM

Quant Time: Nov 25 05:07:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	171269	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	309808	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	314923	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	147934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	104853	52.66	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	105.32%		
35) Dibromofluoromethane	7.55	113	95322	51.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	102.84%		
50) Toluene-d8	10.06	98	380652	50.25	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.50%		
62) 4-Bromofluorobenzene	12.37	95	150778	55.62	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	111.24%		
Target Compounds						
16) Acetone	3.78	43	4755	7.265	ug/l	90
19) Methyl tert-butyl Ether	5.00	73	3532m	0.663	ug/l	
30) Chloroform	7.32	83	2309	0.683	ug/l	97
64) Tetrachloroethene	10.60	164	2078	0.678	ug/l #	75
95) Naphthalene	15.11	128	6603	0.736	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064812.D
Acq On : 24 Nov 2020 15:51
Operator : JC/MD
Sample : L4829-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11

Manual Integrations
APPROVED

MMDadoda
11/25/2020 3:57:24 PM

Quant Time: Nov 25 05:07:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

