

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064825.D
 Acq On : 24 Nov 2020 21:53
 Operator : JC/MD
 Sample : VN1124WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1124WBL02

Quant Time: Nov 25 05:41:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	163130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	292883	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	298196	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	138257	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	101011	53.26	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	106.52%	
35) Dibromofluoromethane	7.55	113	92289	52.66	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	105.32%	
50) Toluene-d8	10.06	98	362009	50.56	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	101.12%	
62) 4-Bromofluorobenzene	12.37	95	142595	55.64	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	111.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064825.D
Acq On : 24 Nov 2020 21:53
Operator : JC/MD
Sample : VN1124WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1124WBL02

Quant Time: Nov 25 05:41:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

