

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112420\
 Data File : VN064829.D
 Acq On : 24 Nov 2020 23:29
 Operator : JC/MD
 Sample : L4798-02DL 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 15BR-20201118DL

Quant Time: Nov 25 06:20:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	170634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	303486	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	312269	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	139185	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	104105	52.48	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.96%	
35) Dibromofluoromethane		7.55	113	96367	53.07	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.14%	
50) Toluene-d8		10.06	98	376617	50.76	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.52%	
62) 4-Bromofluorobenzene		12.37	95	146090	55.02	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.17	62	7162	3.575	ug/l	99
12) 1,1-Dichloroethene	3.70	96	682	0.412	ug/l #	76
21) trans-1,2-Dichloroethene	4.99	96	809	0.440	ug/l #	77
27) cis-1,2-Dichloroethene	6.78	96	112336	54.331	ug/l	90
44) Trichloroethene	8.80	130	141278	55.548	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112420\
Data File : VN064829.D
Acq On : 24 Nov 2020 23:29
Operator : JC/MD
Sample : L4798-02DL 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
15BR-20201118DL

Quant Time: Nov 25 06:20:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

